


管状内径千分尺



3224-2000C

- 符合JJG内径千分尺国家计量检定规程
- 微分头测量范围: 1000-1050mm
- 分度值: 0.01mm
- 精度: $(7+n+L/50)\mu\text{m}$
n= 延长杆数量, L=最大测量长度(mm)
- 硬质合金测量面



型号	测量范围	微分头行程	延长杆
3224-2000C	1000-2000mm	50mm	50mm, 100mm(2个), 200mm, 500mm
3224-3000C	1000-3000mm	50mm	50mm, 100mm(2个), 200mm, 500mm, 1000mm
3224-4000C	1000-4000mm	50mm	50mm, 100mm(2个), 200mm, 500mm, 1000mm(2个)
3224-5000C	1000-5000mm	50mm	50mm, 100mm(2个), 200mm, 500mm, 1000mm(3个)