


管状内径千分尺



- 符合JJG内径千分尺国家计量检定规程
- 微分头测量范围: 100-125mm
- 分度值: 0.01mm
- 精度: $(3+n+L/50)\mu\text{m}$
n= 延长杆数量, L=最大测量长度(mm)
- 配校对规
- 硬质合金测量面



3225-500C

型号	测量范围	微分头行程	延长杆
3225-125C	100-125mm	25mm	—
3225-500C	100-500mm	25mm	25mm, 50mm, 100mm, 200mm
3225-900C	100-900mm	25mm	25mm, 50mm, 100mm, 200mm, 400mm
3225-1300C	100-1300mm	25mm	25mm, 50mm, 100mm, 200mm, 400mm(2个)
3225-1700C	100-1700mm	25mm	25mm, 50mm, 100mm, 200mm, 400mm(3个)
3225-2100C	100-2100mm	25mm	25mm, 50mm, 100mm, 200mm, 400mm(4个)