

注意: 标准杆选配

检验证书

外径千分尺
←INSIZE→PLUS

- 符合JJG外径千分尺国家计量检定规程
- 硬质合金测量面
- 分度值: 0.01mm
- 棘轮套测力
- 可选附件: 标准杆(型号7386)

型号	测量范围	精度	L	b
3701-25A*	0-25mm	4 μ m	37mm	24mm
3701-50A*	25-50mm	4 μ m	62mm	34mm
3701-75A*	50-75mm	5 μ m	87mm	46mm
3701-100A*	75-100mm	5 μ m	112mm	60mm
3701-125A*	100-125mm	6 μ m	139mm	72mm
3701-150A*	125-150mm	6 μ m	164mm	85mm
3701-175A*	150-175mm	7 μ m	190mm	97mm
3701-200A*	175-200mm	7 μ m	215mm	110mm
3701-225A*	200-225mm	8 μ m	240mm	122mm
3701-250A*	225-250mm	8 μ m	264mm	135mm
3701-275A*	250-275mm	9 μ m	290mm	148mm
3701-300A*	275-300mm	9 μ m	315mm	160mm

*带生产厂检验证书



3701-25A

